

	DESCRIPTION		
	The TRH450 is designed for environmental temperature and		
	humidity acquisition, requiring enhanced precision and an extended		
	temperature range. Thanks to its factory-calibrated, linearized,		
	and temperature-compensated digital sensor chip, it is field		
	interchangeable. Its compact aluminum probe also includes extra		
	physical protection for harsh environmental conditions, and an		
	internal filter protects against dust, soot, and other contaminants. Its		

BOBLIST, HIGH PRECISION USB TEMPERATURE AND HUMIDITY SENSOR

thin probe eases integration, even in space-constrained locations.

Г	R	Н	4	5	ľ

APPLICATIONS ○ OEM

- Greenhouse
- \odot Server rooms
- Manufacturing
- Pre-certification
- LIMS integration
- Humidity control
- Scientific research
- Building automation
- Engineering and R&D
- Environmental chamber

INSTALL	ATION TIME		SPECIFIC	ATIONS			
Less thar	10 minutes	Parameter	Cond			Value	Units
		Temperature					
UNIQUE SERIAL NUMBER Each unit is assigned a unique serial number allowing for		Operating range	-40 to 125 Max		Мах	-	°C
		Accuracy	20 to 60°	C	Typ. Max	±0.1 ±0.3	°C
traceabili	ty and certification	Accuracy	-41110 211-1.		Typ. Max.	±0.2 ±0.3	°C
FREE DA	NQ SOFTWARE	Accuracy	60 to 125°C Typ. Max.		Typ. Max.	±0.4 ±0.6	°C
Real-time and loggi	e data visualization ng	Accuracy	-40 to 125°C Typ. Max.		±0.4 ±0.6	°C	
	-	Resolution	Ту	p.		0.015	°C
DATA IN	TECOATION	Repeatability	Ту			0.06	°C
UAIA IN	TEGRATION	Response time	t63%		10	S	
Command-line tools for direct data access and integration		Factory calibrated	Individually ^[2]		yes	-	
		Long-term drift	Max.		< 0.03	°C/yr	
		Relative humidity					
OPTION	2	Operating range ^[3]	Non-conden	sing	-	0 to 100	%RH
∘ Virtua	al COM Port (VCP)	Accuracy	0 to 55 %RH	25°C	Typ. Max	±1.5 ±2	%RH
	nunication protocol nt user calibration	Accuracy	55 to 100 %RH	25°C	Typ. Max	±2 ±3	%RH
mech	anism	Accuracy	0 to 100 %RH	0 to 80°C	Typ. Max.	±2 ±3	%RH
AI SO AI	/AILABLE	Resolution	Тур.		0.01	%RH	
		Hesterisis	25°C		0.8	%RH	
Iraceabil	ity certificates	Factory calibrated	Individually ^[2]		Yes	-	
DE USEU III applications		Long-term drift ^[5]	Typ., -40 to 70°C <		< 0.25	%RH/yr	
		Probe					
Note: While every effort		Operating range	-40 to 125			°C	
	has been made to ensure accuracy in	Cable material	Silicone				
	this publication, no responsibility can be	Cable length	1 (3)			m (ft)	
	accepted for errors or omissions.	First filter material	Anodized aluminum				
Note:	Data may change	Sec. filter material	PTF	E membr	ane		

SPECIFICATIONS					
Parameter	Condition	Value	Units		
Power supply					
Voltage	Powered through a USB port	5	V		
Current consumption	At 5V	≤18	mA		
Mechanical					
Dimensions	See schema below	-	-		
Colour	-	Cyan	-		
Weight (without USB cable)	-	50	g		
Housing and USB cat	ble				
Temperature operating range	_	0 to 70	°C		
Humidity operating range	Non condensing	10 to 90	%RH		
Material	-	ABS	-		
IP rating ^[3]	-	51	-		
System galvanic isolation	-	None	-		
USB cable length	-	1 (3)	m (ft)		
Miscellaneous					
ADC resolution	-	16	bits		
Long-term stability	Maximum	0.03	°C		
Temperature compensated	By the manufacturer	Yes	-		
Lifetime	-	5	years		
Certification(s)					
RoHS	RoHS3	Yes	_		
CE	CE/REACH	Yes			
^[2] Each sensor is individually conditioned by the manufacturer of the semi-					

Each sensor is individually conditioned by the manufacturer of the semiconductor sensor chips in the best stable conditions, and their correction coefficients are recorded for each of them.

[3] If water condensation or splashing is possible, installing the probe pointing down is recommended to reduce the risk of water build-up in the sensor. If water splashing is possible, take extra precautions to protect the sensor and the cable converter. Depending on the application, extra housing may be required.

[5] Typical value for operation in average relative humidity and temperature range. Maximum value is < 0.5 %RH/yr. Higher drift values might occur due to contaminant environments with vaporized solvents, out-gassing tapes, adhesives, packaging materials, etc. For optimal perfomance, keep the unit in a contaminant free (VOCs) and well ventilated area.

FREE DAQ

DATA INTE

OPTIONS

- Virtual C commur
- 3-point mechan


ALSO AVAI

- Warning: wł ca Wł ha Note: en thi res ac on
- Note: Da Data may cnange without notification, and you are strongly advised to obtain copies of the most rec datasheet. recently issued




Efficiency

Particle size \geq 200 nm 99.99 %



AVAILABLE CHANNEL(S) As displayed in our logging software			
CHANNEL ID*	DECRIPTION	ТҮРЕ	NATURE
00	SHT31 Temperature	Temperature	Real
01	SHT31 Relative Humidity	Relative Humidity	Real
02	Dew point	Dew point	Virtual
03	Humidex	Humidex	Virtual
04	Heat index	Heat index	Virtual

* Channel ID as it appears in DracalView. Virtual channel IDs differ in DracalView and dracal-usb-get.



	ORDERING			
PRODUCT(S)				
PART NUMBER	OPTION	DESCRIPTION		
601034	USB-TRH450	Robust, high precision usb temperature and humidity sensor		
608034	USB-TRH450-CAL	Robust, high precision usb temperature and humidity sensor - calibratable		
603034	VCP-TRH450	Robust, high precision usb temperature and humidity sensor - with VCP mode		
605034	VCP-TRH450-CAL	Robust, high precision usb temperature and humidity sensor- calibratable with VCP mode		
TRACEABILITY CERTIFICATE(S)				
NT1WT 1-point temperature certificate for one (1) unit				
	NT2WT 2-point temperature certificate for one (1) unit			
	1-point relative humidity certificate for one (1) unit			
	2-point relative humidity certificate for one (1) unit			
NT3WH	3-point relative humidity certificate for one (1) unit			
NT4WH	4-point relative humidity certificate for one (1) unit			

- CAUTION: Please remember that electromagnetic interference (EMI) may decrease the accuracy of the sensor. Avoid using this device near EMI sources such as motors, high-voltage transformers, and fluorescent tubes.
 - NOTE: Note that this product is not waterproof and requires protection if contact with water is possible.
 - TIP: Avoid installing the sensor in a location where strong vibration is likely to occur. Strong vibrations may cause slight inaccuracies in the reading.
 - TIP: Before using any precision measurement equipment, it is advised to power the unit for at least 15 minutes.

Visit us at: sales@dracal.com

Sales:

General Inquiries:

info@dracal.com

Technical Support:

support@dracal.com

www.dracal.com

Dracal Technologies Inc. 7900 Taschereau Blvd. Building A, suite 204 Brossard, QC, Canada J4X 1C2